

### To disinfect the Microdot<sup>®</sup> Meter:

Read the Microdot<sup>®</sup> Bleach Wipe (EPA Reg. No. 69687-1-88459) label and follow directions for use. Microdot<sup>®</sup> Bleach Wipe is for the exterior surfaces of the blood glucose meter and is not for the use on the needle or the monitor.

1. Always use personal protective equipment as specified on the Microdot<sup>®</sup> Bleach Wipe label.
2. Thoroughly clean gross filth and heavy soil from surface of Microdot<sup>®</sup> Meter to be disinfected.
3. Open Microdot<sup>®</sup> Bleach Wipe pop-up canister. The wipes are pre-saturated with a sodium hypochlorite (bleach) hospital-use solution.
4. Remove a pre-saturated 6" x 6" wipe.
5. Thoroughly wipe the Microdot<sup>®</sup> Meter surface to be disinfected.
6. Allow the Microdot<sup>®</sup> Meter surface to remain thoroughly wet for 5 minutes.
7. Then allow the Microdot<sup>®</sup> Meter surface to air dry.
8. Dispose of wipe in trash after use. Do not flush wipe in the toilet.
9. Dispose of the non-refillable empty canister according to state and local authorities as allowed by the Microdot<sup>®</sup> Bleach Wipe label.

#### DISINFECTION TIP:

In order to maintain 5 minutes of contact time with the sodium hypochlorite, wrap the MD meter in the wipe and place meter face down.



CAMBRIDGE SENSORS USA, LLC

TOLL FREE 1.877.374.4062

[www.microdotcs.com](http://www.microdotcs.com)

